



Calibration Certificate

Customer Details:

Name: UNITED METALS
Address: Limerick Docks
Limerick
Co. Limerick

Post Code: EIRE

Contact Name: David Hegarty
Tel No: 087 7919470

Equipment Details.

Agreement No:	
Location:	WEIGHBRIDGE
Plant Number:	
Manufacturer:	AWTX
Model:	ZM510
Serial No:	164650488
Capacity:	50000kg
Resolution:	20kg
Tolerance:	As Below
Calibration Frequency:	12 Months

Date of Calibration: 19/04/2022

Certificate No: SC19042022/005

Test Weight Details

Weight/Set Number
W010
T163

Certificate Number
A0332
UK210640

Test weights used are traceable to national standards of Mass.

Tel: 0845 900 22 44

Service Centre

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments.

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation	Off Centre Loading Positions	2	3
0kg	0kg	Nil	0kg	Nil		1	4
10000kg	10000kg	Nil	10000kg	Nil	As Found Test		
20000kg	20000kg	Nil	20000kg	Nil			
26580kg	26580kg	Nil	26580kg	Nil			
39640kg	39620kg	-20kg	39620kg	-20kg	Pos.1	10580kg	10520kg
46220kg	46180kg	-40kg	46180kg	-40kg	Pos.2	10580kg	10520kg
					Pos.3	10580kg	10540kg
					Pos.4	10580kg	10540kg
					Definitive Test		
					Pos.1	10580kg	10560kg
					Pos.2	10580kg	10560kg
					Pos.3	10580kg	10560kg
					Pos.4	10580kg	10560kg
Repeatability (x3) tested at:					Tolerances		
Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation	Unless otherwise agreed tolerances will be those permitted by current Weights and Measures regulations for Class III machines as follows		
46220kg	46180kg	-40kg	46180kg	-40kg	0-500divs=+/- 1div		
46220kg	46180kg	-40kg	46180kg	-40kg	>500-2000divs=+/-2divs		
46220kg	46180kg	-40kg	46180kg	-40kg	>2000divs=+/-3divs		

Comments:

WB OPERATOR REPORTED LORRY WEIGHTS 100kg + LIGHT
FOUND LARGE BUILD UP OF DIRT ENTIRE LENGH OF WB NEEDS
TO BE CLEANED OUT WITH POWER WASHERS & RODS FOR SOLIDS

Technician Name: S CARMODY

Next Calibration Due: April 2023

Technician Signature:

Customer Signature: _____



Calibration Certificate

Customer Details:

Name: UNITED METALS
 Address: Eastway Business Park
 Limerick
 Co. Limerick
 Post Code: EIRE
 Contact Name: David Hegarty
 Tel No: 087 7919470

Equipment Details.

Agreement No:	
Location:	ENTRANCE
Plant Number:	
Manufacturer:	AWTX
Model:	E1205
Serial No:	100550344
Capacity:	60000kg
Resolution:	20kg
Tolerance:	As Below
Calibration Frequency:	12 Months

Date of Calibration: 19/04/2022

Certificate No: SC19042022/001

Test Weight Details

Weight/Set Number #VALUE!
 Certificate Number A0332

Test weights used are traceable to national standards of Mass.

Tel: 0845 900 22 44

Service Centre

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments.

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation	Off Centre Loading Positions	2	3
0kg	0kg	Nil	0kg	Nil		1	4
10000kg	10000kg	Nil	10000kg	Nil	As Found Test		
20000kg	20000kg	Nil	20000kg	Nil		Test Point	Reading
26600kg	26600kg	Nil	26600kg	Nil	Pos.1	10600kg	10600kg
39580kg	39580kg	Nil	39580kg	Nil	Pos.2	10600kg	10600kg
46180kg	46180kg	Nil	46180kg	Nil	Pos.3	10600kg	10600kg
					Pos.4	10600kg	10600kg
					Definitive Test		
						Test Point	Reading
					Pos.1	10600kg	10600kg
					Pos.2	10600kg	10600kg
					Pos.3	10600kg	10600kg
					Pos.4	10600kg	10600kg
Repeatability (x3) tested at:					Tolerances		
Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation	Unless otherwise agreed tolerances will be those permitted by current Weights and Measures regulations for Class III machines as follows		
46180kg	46180kg	Nil	46180kg	Nil	0-500divs=+/- 1div		
46180kg	46180kg	Nil	46180kg	Nil	>500-2000divs=+/-2divs		
46180kg	46180kg	Nil	46180kg	Nil	>2000divs=+/-3divs		

Comments:

Technician Name: S CARMODY

Next Calibration Due: April 2023

Technician Signature: *S. Carmody*

Customer Signature: _____



Calibration Certificate

Customer Details:

Name: UNITED METALS
 Address: Ballysimon Rd
 Limerick
 Co. Limerick
 Post Code: EIRE
 Contact Name: David Hegarty
 Tel No: 087 7919470

Equipment Details:

Agreement No:	
Location:	SMALL WB
Plant Number:	
Manufacturer:	AWTX
Model:	E1205
Serial No.:	094050274
Capacity:	30000kg
Resolution:	10kg
Tolerance:	As Below
Calibration Frequency:	12 Months

Date of Calibration: 19/04/2022

Certificate No: SC19042022/003

Test Weight Details

Weight/Set Number
 W010
 T163

Certificate Number
 A0332
 UK210640

Test weights used are traceable to national standards of Mass.

Tel: 0845 900 22 44

Service Centre

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments.

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation	Off Centre Loading Positions			
					2	3		
0kg	0kg	Nil	0kg	Nil	As Found Test			
5000kg	5000kg	Nil	5000kg	Nil				
10000kg	10000kg	Nil	10000kg	Nil	Test Point	Reading	Deviation	
15000kg	15000kg	Nil	15000kg	Nil	Pos.1	10590kg	10590kg	Nil
20000kg	20000kg	Nil	20000kg	Nil	Pos.2	10590kg	10590kg	Nil
26950kg	26590kg	Nil	26590kg	Nil	Pos.3	10590kg	10590kg	Nil
					Pos.4	10590kg	10590kg	Nil
					Definitive Test			
					Test Point	Reading	Deviation	
					Pos.1	10590kg	10590kg	Nil
					Pos.2	10590kg	10590kg	Nil
					Pos.3	10590kg	10590kg	Nil
					Pos.4	10590kg	10590kg	Nil

Repeatability (x3) tested at:

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
20000kg	20000kg	Nil	20000kg	Nil
20000kg	20000kg	Nil	20000kg	Nil
20000kg	20000kg	Nil	20000kg	Nil

Tolerances

Unless otherwise agreed tolerances will be those permitted by current Weights and Measures regulations for Class III machines as follows

0-500divs=+/- 1div
 >500-2000divs=+/-2divs
 >2000divs=+/-3divs

Comments:

Technician Name: S CARMODY

Next Calibration Due: April 2023

Technician Signature: *S. Carmody*

Customer Signature: _____